

The chemistode: A droplet-based microfluidic device for stimulation and recording with high temporal, spatial, and chemical resolution

Delai Chen^{a,1}, Wenbin Du^{a,1}, Ying Liu^{a,1}, Weishan Liu^{a,1}, Andrey Kuznetsov^b, Felipe E. Mendez^b, Louis H. Philipson^b, and Rustem F. Ismagilov^{a,2}

^aDepartment of Chemistry and Institute for Biophysical Dynamics, University of Chicago, 929 East 57th Street, Chicago, IL 60637; and ^bDepartment of Medicine and the Kovler Diabetes Center, University of Chicago, 5841 South Maryland Avenue, Chicago, IL 60637

Edited by George M. Whitesides, Harvard University, Cambridge, MA, and approved September 15, 2008 (received for review August 19, 2008)

Microelectrodes enable localized electrical stimulation and recording, and they have revolutionized our understanding of the spatiotemporal dynamics of systems that generate or respond to electrical signals. However, such comprehensive understanding of systems that rely on molecular signals—e.g., chemical communication in multicellular neural, developmental, or immune systems—remains elusive because of the inability to deliver, capture, and interpret complex chemical information. To overcome this challenge, we developed the “chemistode,” a plug-based microfluidic device that enables stimulation, recording, and analysis of molecular signals with high spatial and temporal resolution. Stimulation with and recording of pulses as short as 50 ms was demonstrated. A pair of chemistodes fabricated by multilayer soft lithography recorded independent signals from 2 locations separated by 15 μm . Like an electrode, the chemistode does not need to be built into an experimental system—it is simply brought into contact with a chemical or biological substrate, and, instead of electrical signals, molecular signals are exchanged. Recorded molecular signals can be injected with additional reagents and analyzed off-line by multiple, independent techniques in parallel (e.g., fluorescence correlation spectroscopy, MALDI-MS, and fluorescence microscopy). When recombined, these analyses provide a time-resolved chemical record of a system’s response to stimulation. Insulin secretion from a single murine islet of Langerhans was measured at a frequency of 0.67 Hz by using the chemistode. This article characterizes and tests the physical principles that govern the operation of the chemistode to enable its application to probing local dynamics of chemically responsive matter in chemistry and biology.

analysis | dispersion | flow | microscale | pulse

This article describes the “chemistode,” a droplet-based microfluidic device for manipulating and observing molecular signals with high spatial and temporal resolution. The microelectrode, voltage-clamp, and patch-clamp techniques (1) enabled stimulation and recording of electrical activity and redox-active molecules with high resolution in both space and time, revolutionizing our understanding of electroactive processes from biochemistry to neuroscience (1–3). Most biological processes, however, are fundamentally chemical rather than electrical, relying on molecular signals to orchestrate events at the correct time and location. Electrochemical approaches are widely used, but not all molecules are electrochemically active, and some electrochemically active molecules are difficult to measure selectively in complex mixtures. The grand challenge this article addresses is that of devising an analogue of the electrode that operates on molecular rather than electrical or electroactive signals.

Why could we not build such a system with today’s technology? Whereas electrical signals travel through wires essentially instantly and with low losses, manipulation and transport of molecules is more challenging. First, a pulse of molecules,

especially of small volume, rapidly disperses when transported through a tube by laminar flow, leading to loss of concentration and time resolution. Loss of molecules from solution by adsorption to surfaces of tubes may also occur. Therefore, methods that rely on laminar flow to transport molecular signals, such as direct sampling (4), push/pull perfusion (5), microdialysis (6), and direct microinjection, have not addressed this grand challenge. In contrast to electrical signals, molecular signals comprise multiple, often unknown, molecular species, requiring the ability to deliver multiple molecular species as stimuli and the ability to analyze a pulse of response molecules by multiple techniques. Advances in optical imaging technology, new probes and tagging methods, and photo-controllable manipulation have enabled observation and manipulation of many known molecular species, but these technologies may be time consuming to develop for each species and difficult to use for multiple or unknown species. “Biology on a chip” microfluidics technologies (7–9) can reduce dispersion by minimizing the distance that molecules are transported by the integration of a biological experiment with a specific analytical method. However, this approach requires the redevelopment and validation of the biological protocols as well as the miniaturization and integration of disparate analytical technologies. Recent advances in microfluidics have used multiphase flow to transport solutions reliably as discrete units without dilution, cross-contamination, or loss of temporal resolution (10–19).

We developed the chemistode, a microfluidic platform that addresses this grand challenge by providing molecular stimulation and recording with high fidelity using plug-based (12) multiphase microfluidics [Fig. 1*A* and supporting information (SI) Fig. S1]. Like the electrode, the chemistode is simply brought into contact with the surface under investigation, e.g., a cell or tissue. Instead of exchanging electrical signals, molecular signals are delivered by and captured in plugs, aqueous droplets nanoliters in volume surrounded by a fluorocarbon carrier fluid. The compartmentalization of these molecular signals eliminates dispersion and loss of sample due to surface adsorption (18).

Operation of the chemistode relies on 9 general steps (Fig. 1*A* and *B*): (i) preparation of an array of aqueous plugs containing an arbitrary sequence of stimuli (20, 21); (ii) delivery of the array of stimulus plugs to a hydrophilic substrate; (iii) coalescence of

Author contributions: D.C., W.D., Y.L., W.L., A.K., L.H.P., and R.F.I. designed research; D.C., W.D., Y.L., W.L., and F.E.M. performed research; D.C., W.D., Y.L., W.L., A.K., L.H.P., and R.F.I. analyzed data; and D.C., W.D., Y.L., W.L., and R.F.I. wrote the paper.

The authors declare no conflict of interest.

This article is a PNAS Direct Submission.

¹D.C., W.D., Y.L., and W.L. contributed equally to this work.

²To whom correspondence should be addressed. E-mail: r-ismagilov@uchicago.edu.

This article contains supporting information online at www.pnas.org/cgi/content/full/0807916105/DCSupplemental.

© 2008 by The National Academy of Sciences of the USA




Fig. 3. An array of chemistodes operates at high spatial resolution. (A) A schematic drawing of the 2-layer chemistode device used for sampling 2 signals, 8-methoxypyrene-1,3,6 trisulfonic acid (MPTS, blue) or fluorescein (green), released through 2 orifices separated by 15 μm (see [S1 Text](#)). (B–E) A plot of fluorescence intensity of the 2 fluorescent signals, observed through green (g) and blue (b) filters, captured and transported by plugs of the chemistode (B and C) and laminar flow (D and E) in the same geometry.

The solutions were buffer (colorless), fluorescein (green), and 8-methoxypyrene-1,3,6 trisulfonic acid (MPTS, blue). We then brought the pair of chemistodes in contact with the wetting layer above the orifices to record ejected pulses, and we detected fluorescence at the tip of the device (site 1) and 7 cm downstream (site 2). The 2-layer chemistode reliably recorded the sequences of pulses at both locations (Fig. 3 B and C) with cross-contamination of $<1\%$ and no loss of intensity during transport (Fig. 3C). The use of plugs was essential—when single-phase laminar flow was used instead, pulses rapidly broadened, overlapped, and decayed (Fig. 3 D and E). In these small channels, Taylor dispersion was less severe, but losses to the walls of channels became pronounced, especially when we tested solutions of proteins. Because the carrier fluid completely encapsulates the aqueous plug and eliminates its contact with the walls, two-phase flow of the chemistode reliably transported even molecules that tended to adsorb to PDMS and Teflon. We do not anticipate partitioning of hydrophobic or amphiphilic molecules from the aqueous plugs into the fluorocarbon carrier fluid, as shown in previous studies (33, 34).

To test the compatibility of the chemistode with off-line, multianalyte measurements by independent methods (steps vi–ix), we used the chemistode to record pulses of a mixture of 4 compounds— CaCl_2 , insulin, glucose, and MPTS as a positive control—each representing a different class of molecules and detectable by a different technique (Fig. 4A). A mixture of these compounds was pulsed through the end of a Y-shaped channel. The chemistode was brought into contact with the wetting layer above the channel. The array of recorded response plugs was split into 4 identical daughter arrays (step vi). Each plug in the first daughter array was injected (step vii) with a fluorescent indicator, fluo-4. Measuring fluorescence of the fluo-4- Ca^{2+} complex of each plug (step viii) in this array detected the presence of Ca^{2+} ions and provided a profile of Ca^{2+} release as a function of time (Fig. S6). In parallel, plugs in the second array were injected with the mixture of an anti-insulin antibody and

labeled insulin for a competitive immunoassay with a 20 nM limit of detection, shown as the baseline in Fig. 4B and Fig. S7. Control experiments indicated that plugs provide an excellent transport and storage medium for insulin for off-line analysis with no losses of insulin due to degradation or adsorption to surfaces (25), in contrast to almost complete loss of insulin from laminar flow in the same Teflon tube. To determine the concentration of insulin, the fraction of labeled insulin that was free or bound to the antibody was detected by using fluorescence correlation spectroscopy (FCS) (35). Plugs in the third array were injected with Girard's reagent T [(carboxymethyl)trimethylammonium chloride hydrazide] and incubated overnight to give a hydrazone derivative of glucose, and the presence of the hydrazone was then detected by MALDI-MS (Fig. S8, a method that can detect unknown molecules. As a control, fluorescence of MPTS was measured in the fourth array. Analysis of insulin data indicated that $>98\%$ of the pulsed signal was recovered by the chemistode. Final recombination of data from all 4 analyses (step ix) showed good alignment among different techniques and with the positive control trace of MPTS (Fig. 4B).

Splitting followed by off-line analysis (Fig. 4A) is an attractive feature of the chemistode that decouples the stimulating and recording experiment from the equipment and expertise that may be required for analysis of nanoliter volumes. We demonstrated decoupling in time: Whereas the chemical signals were recorded on the time scale of seconds, incubation and measurement steps during analysis required >24 h but did not lead to loss of signal or time resolution. We also demonstrated decoupling in location: We performed recording with the chemistode (Fig. 4A) in our laboratory and analyzed plugs for insulin 1 day later by using an FCS instrument located 20 miles away. Control experiments indicated that storage and transportation of tubing containing response plugs did not affect the time resolution or quality of analysis by FCS. Arrays of plugs have been shipped cross-country or frozen and thawed without disruption. The ability to deliver, record with no losses of resolution or concentration, duplicate, store, send by mail, and analyze by multiple techniques space- and time-resolved chemical information encoded as an array of plugs in a tube could dramatically enhance collaborations and use of unique analytical facilities.

Finally, we tested the compatibility of the chemistode with live-cell experiments (Fig. 4C) by using mouse islets of Langerhans, a model system widely studied in the context of diabetes (36–38). Using the chemistode, we stimulated single islets (Fig. 4D) by using a transition from a resting buffer containing 2 mM glucose to a stimulant buffer containing 14 mM glucose. The increase of intracellular $[\text{Ca}^{2+}]_i$ by islets in response to the stimulation was optically monitored (36) by measuring the fluorescence intensity of fluo-4. We observed regularly oscillating $[\text{Ca}^{2+}]_i$ in islets being stimulated with plugs of buffer containing 14 mM glucose for up to 1 h (Fig. S9). Next, by forming recording plugs at a frequency of 0.67 s^{-1} and measuring the insulin concentration in the recording plugs, we determined, every 1.5 s, the rate of insulin secretion of an islet under stimulation with solutions containing 30 mM KCl and 2 mM glucose. We chose these conditions to ensure rapid response dynamics and rapid secretion of insulin. We included a fluorescent dye, Alexa Fluor 594, as a marker in the stimulant plugs. To align the trace of off-line analysis to the $[\text{Ca}^{2+}]_i$ data, we used the transitions of marker intensity between the baseline and the plateau as the references and kept track of the position of every plug in the recording array. The intensity of Alexa Fluor 594 measured by fluorescent microscopy during stimulation agreed well with the intensity measured by off-line analysis of recorded plugs, indicating that temporal information was preserved by plugs. Upon stimulation, after a short ≈ 2 -s delay, the islet displayed the expected response—a sharp increase of $[\text{Ca}^{2+}]_i$, accompanied by a burst of insulin release within ≈ 10 s (Fig. 4E).

Materials and Methods

See *SI Text* for materials, more detailed procedures, and additional data.

Fabrication and Operation of Microfluidic Devices. All microfluidic devices were fabricated by rapid-prototyping soft lithography in PDMS (24). Surfaces of microchannels were made hydrophobic and fluorophilic by silanization (*SI Text*) (25). The chemistode device was fabricated by inserting connecting Teflon microcapillaries into the 2 arms of a V-shaped channel. Arrays of stimulant plugs were generated by using a microfluidic device according to previously reported procedures (20) or a home-built robotic system (*SI Text*). Gastight syringes (series 1700; Hamilton) with removable 27-gauge needles and 30-gauge Teflon tubing (Weico Wire and Cable) were used to load aqueous solutions and carrier fluid. PHD 2000 infusion syringe pumps (Harvard Apparatus) controlled with LabVIEW (National Instruments) programs were used to drive flows.

1. Hamill OP, Marty A, Neher E, Sakmann B, Sigworth FJ (1981) Improved patch-clamp techniques for high-resolution current recording from cells and cell-free membrane patches. *Pflügers Arch* 391:85–100.
2. Hubel DH, Wiesel TN (1959) Receptive fields of single neurones in the cats striate cortex. *J Physiol (London)* 148:574–591.
3. Wightman RM (2006) Probing cellular chemistry in biological systems with microelectrodes. *Science* 311:1570–1574.
4. Kennedy RT, Thompson JE, Vickroy TW (2002) In vivo monitoring of amino acids by direct sampling of brain extracellular fluid at ultralow flow rates and capillary electrophoresis. *J Neurosci Methods* 114:39–49.
5. Cellar NA, Burns ST, Meiners JC, Chen H, Kennedy RT (2005) Microfluidic chip for low-flow push-pull perfusion sampling in vivo with on-line analysis of amino acids. *Anal Chem* 77:7067–7073.
6. Stuart JN, Hummon AB, Sweedler JV (2004) The chemistry of thought: Neurotransmitters in the brain. *Anal Chem* 76:120A–128A.
7. El-Ali J, Sorger PK, Jensen KF (2006) Cells on chips. *Nature* 442:403–411.
8. Atencia J, Beebe DJ (2005) Controlled microfluidic interfaces. *Nature* 437:648–655.
9. Sabounchi P, et al. (2006) Soft-state biomicrofluidic pulse generator for single cell analysis. *Appl Phys Lett* 88:183901.
10. Anna SL, Bontoux N, Stone HA (2003) Formation of dispersions using “flow focusing” in microchannels. *Appl Phys Lett* 82:364–366.
11. Fidalgo LM, Abell C, Huck WTS (2007) Surface-induced droplet fusion in microfluidic devices. *Lab Chip* 7:984–986.
12. Song H, Tice JD, Ismagilov RF (2003) A microfluidic system for controlling reaction networks in time. *Angew Chem Int Ed* 42:768–772.
13. Lorenz RM, Edgar JS, Jeffries GDM, Chiu DT (2006) Microfluidic and optical systems for the on-demand generation and manipulation of single femtoliter-volume aqueous droplets. *Anal Chem* 78:6433–6439.
14. Garstecki P, Fuerstman MJ, Stone HA, Whitesides GM (2006) Formation of droplets and bubbles in a microfluidic T-junction—Scaling and mechanism of break-up. *Lab Chip* 6:437–446.
15. Link DR, et al. (2006) Electric control of droplets in microfluidic devices. *Angew Chem Int Ed* 45:2556–2560.
16. Wang M, Roman GT, Schultz K, Jennings C, Kennedy RT (2008) Improved temporal resolution for in vivo microdialysis by using segmented flow. *Anal Chem* 80:5607–5615.
17. Sahoo HR, Kralj JG, Jensen KF (2007) Multistep continuous-flow microchemical synthesis involving multiple reactions and separations. *Angew Chem Int Ed* 46:5704–5708.
18. Song H, Chen DL, Ismagilov RF (2006) Reactions in droplets in microfluidic channels. *Angew Chem Int Ed* 45:7336–7356.
19. Huebner A, et al. (2008) Microdroplets: A sea of applications? *Lab Chip* 8:1244–1254.
20. Zheng B, Ismagilov RF (2005) A microfluidic approach for screening submicroliter volumes against multiple reagents by using preformed arrays of nanoliter plugs in a three-phase liquid/liquid/gas flow. *Angew Chem Int Ed* 44:2520–2523.
21. Zheng B, Ismagilov RF (2005) A microfluidic approach for screening submicroliter volumes against multiple reagents by using preformed arrays of nanoliter plugs in a three-phase liquid/liquid/gas flow. *Angew Chem* 117:2576–2579.
22. Adamson DN, Mustafi D, Zhang JXJ, Zheng B, Ismagilov RF (2006) Production of arrays of chemically distinct nanoliter plugs via repeated splitting in microfluidic devices. *Lab Chip* 6:1178–1186.
23. Li L, Boedicker JQ, Ismagilov RF (2007) Using a multijunction microfluidic device to inject substrate into an array of preformed plugs without cross-contamination: Comparing theory and experiments. *Anal Chem* 79:2756–2761.
24. Duffy DC, McDonald JC, Schueller OJA, Whitesides GM (1998) Rapid prototyping of microfluidic systems in poly(dimethylsiloxane). *Anal Chem* 70:4974–4984.
25. Roach LS, Song H, Ismagilov RF (2005) Controlling nonspecific protein adsorption in a plug-based microfluidic system by controlling interfacial chemistry using fluorosurfactants. *Anal Chem* 77:785–796.
26. Tice JD, Song H, Lyon AD, Ismagilov RF (2003) Formation of droplets and mixing in multiphase microfluidics at low values of the Reynolds and the capillary numbers. *Langmuir* 19:9127–9133.
27. Adzima BJ, Velankar SS (2006) Pressure drops for droplet flows in microfluidic channels. *J Micromech Microeng* 16:1504–1510.
28. Bibette J, Calderon FL, Poulin P (1999) Emulsions: Basic principles. *Rep Prog Phys* 62:969–1033.
29. Rekvig L, Frenkel D (2007) Molecular simulations of droplet coalescence in oil/water/surfactant systems. *J Chem Phys* 127.
30. Pawar Y, Stebe KJ (1996) Marangoni effects on drop deformation in an extensional flow: The role of surfactant physical chemistry. 1. Insoluble surfactants. *Phys Fluids* 8:1738–1751.
31. Anderson JR, et al. (2000) Fabrication of topologically complex three-dimensional microfluidic systems in PDMS by rapid prototyping. *Anal Chem* 72:3158–3164.
32. Unger MA, Chou HP, Thorsen T, Scherer A, Quake SR (2000) Monolithic microfabricated valves and pumps by multilayer soft lithography. *Science* 288:113–116.
33. Li L, et al. (2006) Nanoliter microfluidic hybrid method for simultaneous screening and optimization validated with crystallization of membrane proteins. *Proc Natl Acad Sci USA* 103:19243–19248.
34. Hatakeyama T, Chen DLL, Ismagilov RF (2006) Microgram-scale testing of reaction conditions in solution using nanoliter plugs in microfluidics with detection by MALDI-MS. *J Am Chem Soc* 128:2518–2519.
35. Haustein E, Schwille P (2007) Fluorescence correlation spectroscopy: Novel variations of an established technique. *Annu Rev Biophys Biomolec Struct* 36:151–169.
36. Rocheleau JV, Walker GM, Head WS, McGuinness OP, Piston DW (2004) Microfluidic glucose stimulation reveals limited coordination of intracellular Ca^{2+} activity oscillations in pancreatic islets. *Proc Natl Acad Sci USA* 101:12899–12903.
37. Ma L, et al. (2004) Direct imaging shows that insulin granule exocytosis occurs by complete vesicle fusion. *Proc Natl Acad Sci USA* 101:9266–9271.
38. Dishinger JF, Kennedy RT (2007) Serial immunoassays in parallel on a microfluidic chip for monitoring hormone secretion from living cells. *Anal Chem* 79:947–954.
39. Matsumoto S, Kuroda Y (2002) Perfluorocarbon for organ preservation before transplantation. *Transplantation* 74:1804–1809.
40. Juncker D, Schmid H, Delamarche E (2005) Multipurpose microfluidic probe. *Nat Mater* 4:622–628.
41. Olofsson J, et al. (2005) A chemical waveform synthesizer. *Proc Natl Acad Sci USA* 102:8097–8102.
42. Takayama S, et al. (2001) Laminar flows—Subcellular positioning of small molecules. *Nature* 411:1016.
43. Chen X, Kis A, Zettl A, Bertozzi CR (2007) A cell nanoinjector based on carbon nanotubes. *Proc Natl Acad Sci USA* 104:8218–8222.
44. Beebe DJ, et al. (2000) Functional hydrogel structures for autonomous flow control inside microfluidic channels. *Nature* 404:588–590.
45. Kost J, Langer R (2001) Responsive polymeric delivery systems. *Adv Drug Deliv Rev* 46:125–148.
46. Roelfaers MJB, et al. (2006) Spatially resolved observation of crystal-face-dependent catalysis by single turnover counting. *Nature* 439:572–575.
47. Kastrup CJ, Runyon MK, Shen F, Ismagilov RF (2006) Modular chemical mechanism predicts spatiotemporal dynamics of initiation in the complex network of hemostasis. *Proc Natl Acad Sci USA* 103:15747–15752.
48. Huse M, Lillemeier BF, Kuhns MS, Chen DS, Davis MM (2006) T cells use two directionally distinct pathways for cytokine secretion. *Nat Immunol* 7:247–255.
49. Shea SD, Margoliash D (2003) Basal forebrain cholinergic modulation of auditory activity in the zebra finch song system. *Neuron* 40:1213–1226.
50. Ismagilov RF, Maharbiz MM (2007) Can we build synthetic, multicellular systems by controlling developmental signaling in space and time? *Curr Opin Chem Biol* 11:604–611.

Supporting Information. The following procedures are described in *SI Text*: characterization of the temporal and spatial resolution of stimulation and recording by using the chemistode with fluorescent dyes; analysis of insulin, Ca^{2+} , glucose, and MPTS; and the monitoring of insulin secretion from single mouse islets.

ACKNOWLEDGMENTS. We thank Vytas Bindokas for help with confocal microscopy and Jessica M. Price for contributions in writing and editing this manuscript. This work was supported by National Institutes of Health (NIH) Director’s Pioneer Award 1DP1OD003584 and National Science Foundation (NSF) Collaborative Research in Chemistry Grant CHE-0526693 (to R.F.I.), NIH Grants DK48494, DK63493, and DK20595 (to L.H.P.) and the Blum-Kovler Foundation (L.H.P.). R.F.I. is a Cottrell Scholar of Research Corporation and a Camille Dreyfus Teacher-Scholar. A part of this work was performed at the Materials Research Science and Engineering Center microfluidic facility funded by the NSF.